

NAME OF PRODUCT: CORA-CAINE® Denture Analgesic Adhesive SDS DATE: 06/19/2014

SECTION 1: IDENTIFICATION

PRODUCT NAME: CORA-CAINE Denture Analgesic Adhesive

PRODUCT CODES: 16621, 16624
IDENTIFIED USES: Dentistry
USES ADVISED AGAINST: Non-dental use

MANUFACTURER: Harry J. Bosworth Company

ADDRESS: 7227 North Hamlin Avenue, Skokie, Illinois 60076-3999, USA

 TELEPHONE:
 847-679-3400

 FAX:
 847-679-2080

 EMAIL:
 hjbinfo@bosworth.com

 EMERGENCY PHONE:
 800-535-5053 (US and COMMON)

800-535-5053 (US and Canada) 352-323-3500 (International)

SECTION 2: HAZARDS IDENTIFICATION

CLASSIFICATION: Acute aquatic toxicity (Category 3)
Chronic aquatic toxicity (Category 3)

FDA regulated device - exempt from Regulation (US) 29 CFR 1910.1200.

PICTOGRAM: None

LABELING:

SIGNAL WORD: None

HAZARD STATEMENTS: H412 Harmful to aquatic life with long lasting effects.

PRECAUTIONARY STATEMENTS: P201 Obtain special instructions before use.

P202 Do not handle until all safety precautions have been read and understood.

FILE NO.: SDS21

P234 Keep only in original container. P235+P410 Keep cool. Protect from sunlight.

P261 Avoid breathing dust/fume/gas/mist/vapors/spray.

P264 Wash thoroughly after handling.

P270 Do not eat, drink or smoke when using this product.

P273 Avoid release to the environment.

P280 Wear protective gloves/ protective clothing/ eye protection/ face protection.

P301+P330 IF SWALLOWED: Rinse mouth.

P302+P352 IF ON SKIN: Wash with plenty of soap and water.

P304+P340 IF INHALED: Remove victim to fresh air and keep at rest in a position comfortable for

breathing.

P305+P351+P338 IF IN EYES: Rinse cautiously with water for several minutes. Remove contact lenses, if

present and easy to do. Continue rinsing.

P308+P313 IF exposed or concerned: Get medical advice/attention.
P332+P313 If skin irritation occurs: Get medical advice/attention.
P337+P313 If eye irritation persists: Get medical advice/attention.

P391 Collect spillage.
P402 Store in a dry place.

P403+P233 Store in a well-ventilated place. Keep container tightly closed.
P501 Dispose of contents/container to an approved waste disposal plant.

SECTION 3: COMPOSITION/INFORMATION ON INGREDIENTS

COMPONENT	CAS NO.	%WT	OSHA PEL - TWA	ACGIH TLV - TWA	CLASSIFICATION
Benzocaine	94-09-7	14.5-17.7	N/E	N/E	Aquatic Acute 2; Aquatic Chronic 2; H411

For full text of H-statements mentioned in this section, see section 16.

SECTION 4: FIRST-AID MEASURES

INHALATION: Move person into fresh air. If not breathing, give artificial respiration. If symptoms persist, get medical

attention.

SKIN: Wash off with soap and plenty of water. Get medical attention if irritation develops and persists.

EYE: Rinse thoroughly with plenty of water for at least 15 minutes. Get medical attention if irritation develops and

persists.

INGESTION: Never give anything by mouth to an unconscious person. Rinse mouth with water. Consult a physician.

PREPARED BY: SS PAGE 1 OF 4



NAME OF PRODUCT: CORA-CAINE® Denture Analgesic Adhesive

SECTION 5: FIRE-FIGHTING MEASURES

SUITABLE EXTINGUISHING MEDIA: Water spray, alcohol-resistant foam, dry chemical or carbon dioxide.

HAZARDOUS DECOMPOSITION PRODUCTS: Oxides of carbon and nitrogen.

SPECIAL HAZARDS: No further relevant information available.

ADVICE FOR FIREFIGHTERS: Wear self contained breathing apparatus for firefighting if necessary.

SECTION 6: ACCIDENTAL RELEASE MEASURES

PERSONAL PRECAUTIONS: Use personal protective equipment. Avoid breathing vapors, mist or gas. Ensure adequate ventilation.

Evacuate personnel to safe areas.

ENVIRONMENTAL PRECAUTIONS: Prevent further leakage or spillage if safe to do so. Do not let product enter drains. Discharge into the

environment must be avoided.

CONTAINMENT AND CLEANUP: Absorb with liquid-binding material (sand, diatomite, acid binders, universal binders, sawdust). Place in

suitable, closed container for disposal according to official regulations.

SECTION 7: HANDLING AND STORAGE

PRECAUTIONS FOR SAFE HANDLING: Product is intended for dental use only. Handling of this product should be by trained dental healthcare

professionals only. Observe normal care for working with chemicals. Avoid contact with the eyes and skin.

FILE NO.: SDS21

SDS DATE: 06/19/2014

Avoid inhalation of vapor or mist. Keep away from foodstuffs, beverages and animal feed.

CONDITIONS FOR SAFE STORAGE: Store only in the original package. Keep container tightly sealed in a dry and well-ventilated place. Protect

from heat and direct sunlight. Store away from food and beverages.

SECTION 8: EXPOSURE CONTROLS/PERSONAL PROTECTION

ENGINEERING CONTROLS: Handle in accordance with good industrial hygiene and safety practice. Use with adequate general or local

exhaust ventilation to maintain exposure levels below the occupational exposure limits. Local exhaust ventilation is preferred since it prevents contamination dispersion into the work area by controlling it at its source. Provide eyewash and safety shower if contact or splash hazard exists. Wash hands before breaks and

at the end of work.

EYE/FACE PROTECTION: Safety glasses.

SKIN PROTECTION: Glove material impermeable and resistant to the product.

BODY PROTECTION: Protective work clothing.

RESPIRATORY PROTECTION: NIOSH (US) or CEN (EU) approved respirators and components.

ENVIRONMENTAL EXPOSURE: Prevent further leakage or spillage if safe to do so. Do not let product enter drains. Discharge into the

environment must be avoided.

SECTION 9: PHYSICAL AND CHEMICAL PROPERTIES

APPEARANCE/COLOR: Salmon-colored paste

ODOR: None

 BOILING POINT:
 194°F (90°C)

 FLASH POINT:
 > 200°F (> 93°C)

 WATER SOLUBILITY:
 Limited

SECTION 10: STABILITY AND REACTIVITY

CHEMICAL STABILITY: Stable under recommended storage conditions.

HAZARDOUS REACTIONS: Will not occur.

CONDITIONS TO AVOID: Heat.

INCOMPATIBLE MATERIALS: Strong oxidizing agents. **HAZARDOUS DECOMPOSITION PRODUCTS:** Oxides of carbon and nitrogen.

SECTION 11: TOXICOLOGICAL INFORMATION

POTENTIAL HEALTH EFFECTS: EYES: May cause eye irritation with redness, tearing and stinging. May cause mild anesthetizing effect.

<u>SKIN:</u> May cause skin irritation with redness, swelling and itching. May cause mild anesthetizing effect. <u>INGESTION:</u> May cause mucous membrane and gastrointestinal irritation with nausea, vomiting and diarrhea.

May cause anesthetizing effect.

PREPARED BY: SS PAGE 2 OF 4



NAME OF PRODUCT: CORA-CAINE® Denture Analgesic Adhesive SDS DATE: 06/19/2014

INHALATION: No health effects expected. Vapors generated at high temperature may result in respiratory

FILE NO.: SDS21

track irritation.

CARCINOGENICITY: OSHA: No component of this product present at levels greater than or equal to 0.1% is identified as a

carcinogen or possible carcinogen by OSHA.

ACGIH: No component of this product present at levels greater than or equal to 0.1% is identified as a

carcinogen or potential carcinogen by ACGIH.

NTP: No component of this product present at levels greater than or equal to 0.1% is identified as a known or

anticipated carcinogen by NTP.

IARC: No component of this product present at levels greater than or equal to 0.1% is identified as probable,

possible or confirmed human carcinogen by IARC.

SECTION 12: ECOLOGICAL INFORMATION

ADVERSE EFFECTS: An environmental hazard cannot be excluded in the event of unprofessional handling or disposal. Harmful to

aquatic life with long lasting effects. Avoid release to the environment.

SECTION 13: DISPOSAL CONSIDERATIONS

PRODUCT: Offer surplus and non-recyclable solutions to a licensed disposal company. Must not be disposed of together

with household garbage. Do not allow product to reach sewage system. Disposal must be made according to

official regulations.

CONTAMINATED PACKAGING: Dispose of as unused product.

SECTION 14: TRANSPORT INFORMATION

UN NUMBER: N/A
PROPER SHIPPING NAME: N/A
HAZARD CLASS: N/A
PACKING GROUP: N/A
LABEL STATEMENT: N/A

SECTION 15: REGULATORY INFORMATION

US FEDERAL REGULATIONS

TSCA This product is an FDA regulated device and not subject to TSCA regulations.

CERCLA This product is an FDA regulated device and not subject to reporting requirements. There may be specific

reporting requirements at the local, regional, or state level.

SARA 313 TOXIC CHEMICALS The following components are subject to reporting levels established by SARA Title III, Section 313 (40 CFR

372): None

SARA 311/312 HAZARDS This product is an FDA regulated device and not subject to reporting requirements.

INTERNATIONAL REGULATIONS

CANADIAN ENVIRONMENTAL

PROTECTION ACT: This product is a medical device and not subject to chemical notification requirements.

EUROPEAN INVENTORY OF EXISTING COMMERCIAL CHEMICAL SUBSTANCES

(EINECS): This product is a medical device and not subject to chemical notification requirements.

SECTION 16: OTHER INFORMATION

FULL TEXT OF H STATEMENTS REFERRED TO UNDER SECTION 3

Aquatic Acute Acute aquatic toxicity
Aquatic Chronic Chronic aquatic toxicity

H411 Toxic to aquatic life with long lasting effects.

NFPA RATING

Health Hazard 0

PREPARED BY: SS PAGE 3 OF 4



NAME OF PRODUCT: CORA-CAINE® Denture Analgesic Adhesive

Fire Hazard Reactivity Hazard

PREPARATION INFORMATION: This SDS was prepared in accordance with the requirements of the OSHA Hazard Communication Standard (29 CFR 1910.1200) and is to be used only for this product.

DISCLAIMER: To the best of our knowledge, the information contained herein is accurate. However, The Harry J. Bosworth Company does not assume any liability whatsoever for the accuracy or completeness of the information contained herein. Final determination of suitability of any material is the sole responsibility of the user. All materials may present unknown health hazards and should be used with caution. Although certain hazards are described herein, we cannot guarantee that these are the only hazards that exist.

FILE NO.: SDS21

SDS DATE: 06/19/2014

PREPARED BY: SS PAGE 4 OF 4